

Order	The Project Name		Quality technical requirements	
1	Glove surface type		No powder in fingers	
2	material		Nitrile content>95%	
3	Check level and receiving quality limit			
	Size		Checking level	AQL
	(Length、Width、Thickness)		S-2	1.5-2.5
3.1	Length	S	≥220mm	
		M	≥240mm	
		L		
		XL		
3.2	Single layer thickness	Finger	>4mil(0.102mm)	
		Palm	0.08mm±0.01mm	
3.3	Width	S	85mm±5mm	
		M	95mm±5mm	
		L	105mm±5mm	
		XL	115mm±5mm	
4	Impervious		G-1	2.5
5	Breaking force, Elongation at break (Before and after ageing)		S-2	4
6	Single gram		(S)3.5g、(M)4g、(L)4.5g、(XL)5g	
7	Product color		Light blue (Sky blue)	
8	Elongation at break	Before ageing	≥500%	
		After ageing	≥400%	
9	Breaking force		≥7.0N	
10	Sterilization		Non-Sterilization	
11	Packaging	Box	100pcs/box	

		Cardboard	1000pcs/carton
12	Certification standards		<b>ASTM - D6319 Physical testing standards</b>